



# EC type-approval Certificate

Number **T6244** revision 10  
Project number SO13200966  
Page 1 of 1

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to  
conformity modules mentioned in article 9 of Directive 2009/23/EC, after  
having established that the measuring instrument meets the applicable  
requirements of Directive 2009/23/EC, to:

Manufacturer Rinstrum Pty Ltd.  
41 Success Street  
Acacia Ridge, QLD, 4110  
Australia

Measuring instrument **A Non-automatic weighing instrument**  
Type : R300 series

Further properties are described in the annex:  
– Description T6244 revision 10.

Valid until 22 May 2023

Remark This revision replaces the earlier versions.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
4 October 2013

  
C. Oosterman  
Head Certification Board

**NMI Certin B.V.**  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision  
that no liability is accepted and that the  
applicant shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified  
Body can be verified at [http://  
ec.europa.eu/enterprise/newapproach/nando/](http://ec.europa.eu/enterprise/newapproach/nando/)

Parties concerned can lodge  
objection against this decision,  
within six weeks after the date of  
submission, to the general manager  
of NMI (see [www.nmi.nl](http://www.nmi.nl)).

Reproduction of the complete  
document only is permitted.



## 1 General information about the non-automatic weighing instrument

All properties of the non-automatic weighing instrument, whether mentioned or not, shall not be in conflict with the legislation.

### 1.1 Essential parts

Indicator:

Manufacturer	Type	Certificate number
Rinstrum	R300 series	TC6242

Any load cell(s) may be used under this EC type-approval certificate for instruments as described in WELMEC 2.4 Issue 2, provided the following conditions are met:

- There is a respective test certificate (EN45501), a parts certificate (EN45501) or an OIML Certificate of Conformity (R 60) issued for the load cell by a Notified Body responsible for type examination under Directive 2009/23/EC.
- The certificate contains the load cell types and the necessary load cell data required for the manufacturer's declaration of compatibility of modules (WELMEC 2 Issue 5 Section 11), and any particular installation requirements. A load cell marked **NH** is allowed only if humidity testing to EN45501 has been conducted on this load cell.
- The compatibility of load cells and indicator is established by the manufacturer by means of the compatibility of modules form, contained in the above WELMEC 2 document, at the time of EC verification or declaration of EC conformity of type.
- The load transmission must conform to one of the examples shown in WELMEC 2.4 Issue 2.

### 1.2 Essential characteristics

Accuracy class	III or IIII
Maximum number of scale intervals	$n \leq$ the number of scale intervals mentioned in the certificates involved

Further essential characteristics are described in the certificates involved.

### 1.3 Essential shapes

The descriptive markings plate is secured against removal by sealing or will be destroyed when removed and contains the information as required by Directive 2009/23/EC annex IV.

## 1.4 Conditional parts

The non-automatic weighing instrument may be equipped with peripheral equipment which is used for the applications listed in article 1(2)(a) of Directive 2009/23/EC, provided that the peripheral equipment is certified to be connected to an EC type-approved non-automatic weighing instrument by a Notified Body responsible for type examination under Directive 2009/23/EC, or that the equipment and the use of the equipment comply with the requirements of WELMEC 2.5 Issue 2 Section 2.2.

The non-automatic weighing instrument is fitted with a levelling device and a level indicator, unless the instrument is installed in a fixed position. The level indicator has a sensitivity of at least 2 mm for a tilt of 2/1000.

## 1.5 Conditional characteristics

The conditional characteristics are described in the certificates involved.

## 1.6 Non-essential parts

The non-automatic weighing instrument may be connected to non-essential devices, for example but not limited to bar code readers, foot switches, second display's and cash drawers, provided that:

- They do not present primary data used for purposes mentioned in article 1(2)(a) of Directive 2009/23/EC unless the "preliminary observations" in Annex 1 of this directive is satisfied;
- They do not lead to an instrument having other essential characteristics than those fixed by this type-approval document.

Other non-essential parts:

- AC/DC-adapter;
- External power supply.

## 2 Seals

To secure components that may not be dismantled or adjusted by the user, the non-automatic weighing instrument has to be secured in a suitable manner on the locations indicated in the certificates involved.

The connecting cable of the load cell or the junction box is provided with possibility to seal.

## 3 Conditions for conformity assessment

The marks, facilities for the marks and the inscriptions on the non-automatic weighing instrument fulfill the requirements of article 1 of Annex IV of Directive 2009/23/EC.

The compatibility of load cells and indicator is established by the manufacturer by means of the compatibility of modules form, contained in WELMEC 2 Issue 5 Section 11, at the time of EC verification or declaration of EC conformity of type.